



Global Harmonization Working Party
Towards Medical Device Harmonization

FINAL DOCUMENT

Title: GHWP UDI Rule

Authoring Group: Working Group 9 UDI & Nomenclature

Date: [25 November 2023]

Ms. Zhou Wenwen
Chair, WG 9

Ms. Lilo Li
Co-Chair, WG 9

Dr. Li Yi
Secretariat, WG 9

Table of Contents

Chapter I: General Provisions.....	
Error! Bookmark not defined.	
Chapter II: UDI.....	4
Chapter III: UDI carrier.....	5
Chapter IV: UDID.....	6
Chapter V: Supplementary Provisions.....	6

1
2 **GHWP UDI Rule**
3
4

5 **Chapter I: General Provisions**
6

7 The Unique Device Identification System (UDI system) is intended to provide a single,
8 globally harmonized system to adequately identify medical devices through
9 distribution and use. It is critical to note that the maximum benefits of a UDI system
10 can only accrue if all stakeholders, from the manufacturer to healthcare providers and
11 patients and Regulators, use UDI throughout their workflow systems.
12

13 Every medical device needs to be identified by a UDI, unless it is exempted. The
14 regulatory authority of the UDI System shall specify harmonized exemptions for
15 certain devices such as investigational devices and custom made devices from UDI
16 requirements

17 A UDI system includes a Unique Device Identifier (UDI), a UDI carrier, and a UDI
18 Database (UDID).
19

20 – The UDI is a series of numeric or alphanumeric characters that is created
21 through a globally accepted device identification and coding standard.
22

23 – The UDI Carrier is the means to convey the UDI by using AIDC and, if applicable,
24 its HRI.

25 *Note:* Carriers can include, for example, a 1D/linear bar code, a 2D/Matrix bar
26 code, or an RFID system
27

28 – The UDID contains identifying information and other elements associated with
29 the specific medical device.
30

31 1. The regulatory authorities that establish a UDI system are responsible for
32 establishing a standardized UDI system to meet local regulatory requirements and
33 to develop and maintain a local publicly available UDID that is capable of linking,
34 to the extent possible, to other regulatory authority UDIDs. It is recognized that
35 local specificities and regulations could impact certain aspects of UDI
36 implementation.
37

38 Manufacturers are responsible for understanding both regulatory and issuing

39 agency/entity requirements or standards to accurately assign and place the UDI in
40 human readable and AIDC format on the label or on the device itself and on all
41 higher levels of device packaging, as appropriate. Manufacturers are also
42 responsible for the initial submission of, and updates to, the information in the
43 UDID.

44

45 It is recommended that all stakeholders related to UDI actively use it throughout
46 their workflow systems.

47

48 **Chapter II: UDI**

49

50 2. The UDI contains two parts: device identifier (UDI-DI) and production identifier
51 (UDI-PI).

52

53 The UDI-DI is a unique numeric or alphanumeric code specific to a model of
54 medical device and that is also used as the "access key" to information stored in a
55 UDID.

56

57 The UDI-PI is a numeric or alphanumeric code that identifies the unit of device
58 production. The different types of Production Identifier(s) include serial number,
59 lot/batch number, Software as a Medical Device (SaMD) version and
60 manufacturing and/or expiration date.

61

62 A new UDI-DI is required whenever there is a change that could lead to
63 misidentification of the medical device and/or ambiguity in its traceability. Any
64 change of one of the following UDID data elements determines the need for a new
65 UDI-DI:

66

a.Brand Name,

67

b.Device version or model,

68

c.Clinical Size (including Volume, Length, Gauge, Diameter),

69

d.Labeled as single use,

70

e.Packaged sterile,

71

f.Need for sterilization before use,

72

g.Quantity of devices provided in a package,

73

h.Critical warnings or contraindications: e.g. containing latex or DEHP.

74

75 A UDI-DI allocated to a particular medical device should never be reused. Devices
76 that have been withdrawn from the market and are reintroduced may use the

77 original UDI-DI if they are reintroduced without any modifications or changes
78 which require a new UDI-DI.

79

80 3. Manufacturers are responsible for following both regulatory and issuing
81 agency/entity requirements or standards to accurately assign the UDI to the device
82 itself or to the packaging level of the device so that it can be adequately identified
83 through its distribution.

84

85 The UoU UDI-DI is an unmarked identifier assigned to an individual medical
86 device when a UDI is not labeled on the individual device at the level of its unit of
87 use. Its purpose is to provide a UDI-DI to identify a device used on a patient when
88 a UDI-DI does not appear on the label of the device.

89

90 4. An Issuing Agency/Entity is an organization accredited by a regulatory authority to
91 operate a system for the issuance of UDIs. An Issuing Agency/Entity shall meet the
92 following criteria:

93

94 (a) its system for the assignment of UDIs is adequate to identify a device
95 throughout its distribution and use in accordance with the requirements of the
96 regulatory authority and conforms to the relevant international standards;

97 (b) the entity gives access to its system for the assignment of UDIs to all interested
98 users in accordance with a single set of consistent, fair and reasonable terms and
99 conditions;

100 (c) it makes available to the regulatory authority, upon request, information
101 concerning its system for the assignment of UDIs; and

102 (d) it meets the relevant requirements for data security of the local regulatory
103 authority.

104

105 **Chapter III: UDI Carrier**

106

107 5. A UDI Carrier includes both AIDC and HRI formats. If there are significant
108 constraints limiting the use of both on the label, the AIDC format shall be favored.
109 No particular AIDC methods should be required by a regulatory authority. Each
110 regulatory authority may have specific requirements for particular AIDC methods.
111 The HRI format shall follow the rules of the UDI code issuing agency/entity.

112

113 Commonly used AIDC technologies in the medical device industry include
114 1D/linear bar codes, 2D/Matrix bar codes, and RFID. If linear bar codes are used,

115 the UDI-DI and UDI-PI can be concatenated or non-concatenated. Where RFID is
116 used, a linear or 2D bar code shall also be provided on the label.

117

118 6. The manufacturer is responsible for placing the UDI Carrier on the label or on the
119 device itself, and on all higher levels of device packaging, not including shipping
120 containers. The UDI Carrier should be readable during normal use and throughout
121 intended life of the medical device.

122

123 **Chapter IV: UDID**

124

125 7. Regulatory authorities are responsible for developing the UDID in their jurisdiction
126 based upon local policy requirements. However, locally specific data elements
127 should be kept to a minimum. The UDID should follow a globally harmonized
128 approach so that healthcare professionals and patients will have access to a single
129 consistent, and complete source of information about a medical device and its key
130 attributes. The UDID shall be accessible to the public free of charge. Each
131 regulatory authority may have specific requirements for data elements published
132 for public access.

133

134 8. The manufacturer is responsible for the submission of identifying information and
135 other medical device data elements in the UDID before the product launching to
136 the market.

137

138 Manufacturers should update the relevant UDID record before the product
139 launching to the market when a change is made to an element that does not
140 require a new UDI-DI.

141

142 **Chapter V: Supplementary Provisions**

143

144 9. The use of a UDI System will facilitate and simplify the documentation of medical
145 device use in various patient records including traditional as well as electronic
146 health records and registries. A UDI system should also enable linkages of medical
147 device information across various systems and across geographies. These
148 applications of UDI could help to identify medical device problems and enhance
149 data analysis.

150

151 The global use of a UDI will facilitate traceability throughout distribution. In order
152 to achieve traceability, it is necessary to involve all stakeholders in the capture and
153 recording of the UDI (UDI-DI + UDI-PI) throughout distribution and use.

154

155

156 10. The definition of the following terms in this rule:

157 *Automatic Identification and Data Capture (AIDC)*

158 A technology used to automatically capture data. AIDC technologies include bar
159 codes, smart cards, biometrics and RFID.

160

161 *Human Readable Interpretation (HRI)*

162 Human Readable Interpretation is a legible interpretation of the data characters
163 encoded in the UDI Carrier.

164

165 *Shipping containers*

166 Shipping container is a container where the traceability is controlled by a process
167 specific to logistics systems.

168

169 11. A risk-based approach is essential to facilitate an effective implementation of UDI
170 system. Implementation should be phased in over a period of years based on
171 product risk classes, starting with the highest risk class, to reduce the burden of
172 implementation.

173

174 **References**

175

176 IMDRF/UDI WG/N7Final: 2013 - UDI Guidance: Unique Device Identification (UDI) of
177 Medical Devices

178 IMDRF/UDI WG/N48 FINAL: 2019- Unique Device Identification system (UDI system)
179 Application Guide

180 US, EU, and China, UDI rule

181

182 **Abbreviations**

183

184 Automatic Identification and Data Capture (AIDC)

185 Device Identifier (UDI-DI)

186 Human Readable Interpretation (HRI)

187 Production Identifier (UDI-PI)

188 Software as a Medical Device (SaMD)

189 Unique Device Identification system (UDI system)

190 Unique Device Identifier (UDI)

191 Unit of Use (UoU)

192

193

194 ----- END OF DOCUMENT -----

195

196

197

198

199

200

201

202